



Test Certificate

Part 250 mL Cylinder Holder, complete

Part Number 10757



			USP / Ph.Eur. Tolerances
Parameter	Cylinder Holder	Mass	450 g ± 10 g

The accuracy of the above mentioned part parameters have been examined. All relevant measurements are within tolerance and the factory acceptance test has passed. The parameters are in conformance with USP and Ph. Eur. Pharmacopeia requirements (at date of approval, see below).

Aesch, 2021-11-09

Bruno Balli
SOTAX Quality Management
SOTAX AG | Nordring 1 | 4147 Aesch | Switzerland